



वै0औ0अ0प0 - केन्द्रीय औषधीय एवं संगंध पौधा संस्थान

(वैज्ञानिक तथा औद्योगिक अनुसंधान परिषद)

पोस्ट ऑफिस- सीमैप, लखनऊ-226015, भारत

CSIR-CENTRAL INSTITUTE OF MEDICINAL & AROMATIC PLANTS

(Council of Scientific & Industrial Research)

P.O. CIMAP, Lucknow - 226 015, (U.P.) INDIA

CORRIGENDUM

File No. 113(1)/2018-Pur.


Date 04.12.2018

In reference to our Tender No. 113(1)/2018-Pur. T-25 dated 26.10.2018 having tender ID 2018_CSIR_17160_1, following amendments have been made-

Sr. No.	Description	Existing	Read As
01.	Point No. 9 Refrigerated Circulating Temperature Bath	Refrigerated Circulating Temperature Bath Having suitable bath and Cooling/ heating capacities used at different points in the system viz. for circulating the coolant in pre cooling heat exchanger, in high pressure CO2 pump suction and in CO2 recycle system at 5°C and at separators with temperature Range: ambient (-20°C) to 120° C	Refrigerated Circulating Temperature Bath Having suitable bath and Cooling heating capacities used at different points in the system viz. for circulating the coolant in pre cooling heat exchanger, in high pressure CO2 pump suction and in CO2 recycle system at 5°C and at separators with temperature Range: ambient to 120° C
02.	Point No. 14 UPS	15 KVA Online UPS 0.8-0.9 Power Factor, along with Power Bank Batteries for at least 1 Hour Back Up on load with 02 years Warranty	5 KVA Online UPS, 0.8-0.9 Power Factor, along with Power Bank Batteries for at least 1 Hour Back Up on load with 02 years Warranty
03.	Date of Submission of Tender	04.12.2018 at 03:00 P.M.	11.12.2018 at 03:00 P.M.
04.	Date of Opening of Tender	05.12.2018 at 03:00 P.M.	12.12.2018 at 03: P.M

Others terms & conditions of the tender will remain unchanged.

Yours' faithfully,


Stores & Purchase Officer
Ph : 2718613